

AD _____

GRANT NUMBER: DAMD17-94-J-4304

TITLE: Postdoctoral Fellowship Study of X-Ray Optical Mammograph

PRINCIPAL INVESTIGATOR: Dr. Carolyn A. MacDonald
Dr. Carmen C. Abreu

CONTRACTING ORGANIZATION: State University of New York
University at Albany
Albany, New York 12222

REPORT DATE: October 1995

TYPE OF REPORT: Annual

PREPARED FOR: U.S. Army Medical Research and Materiel Command
Fort Detrick, Maryland 21702-5012

DISTRIBUTION STATEMENT: Approved for public release;
distribution unlimited

The views, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy or decision unless so designated by other documentation.

19960123 120

DTIC QUALITY INSPECTED 1

REPORT DOCUMENTATION PAGE			Form Approved OMB No. 0704-0188	
Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.				
1. AGENCY USE ONLY (Leave blank)	2. REPORT DATE October 1995	3. REPORT TYPE AND DATES COVERED Annual 30 Sep 94 - 29 Sep 95		
4. TITLE AND SUBTITLE Postdoctoral Fellowship Study of X-Ray Optical Mammograph		5. FUNDING NUMBERS DAMD17-94-J-4304		
6. AUTHOR(S) Dr. Carolyn A. MacDonald and Dr. Carmen C. Abreu				
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) State University of New York University at Albany Albany, New York 12222		8. PERFORMING ORGANIZATION REPORT NUMBER		
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) U.S. Army Medical Research and Materiel Command Fort Detrick, Maryland 21702-5012		10. SPONSORING/MONITORING AGENCY REPORT NUMBER		
11. SUPPLEMENTARY NOTES				
12a. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release; distribution unlimited		12b. DISTRIBUTION CODE		
13. ABSTRACT (Maximum 200 words) A number of feasibility measurements of the x-ray imaging properties of a capillary prototype optic for digital mammography were performed. The capillary prototype optic had input and output diameters of 4.15 mm and 7.5 mm, respectively. A mammographic scanning system was constructed to allow imaging of a larger field for some of the measurements. The optic showed good primary transmission (46%) at mammography energies. Images of a lucite contrast detail phantom, using a lead blocker, showed that the optic provides an increase in image contrast with a contrast improvement of 1.72 and a reduction of the scatter fraction to 0.018. The limiting resolution for the phosphor plate system (CR) was measured through the modulation transfer function (MTF). The 5% MTF for the stationary and scanned optic was 9 lp/mm and 8.4 lp/mm, respectively, compared with 5.4 lp/mm without the optics using a 0.3 mm focal spot and optimal magnification.				
14. SUBJECT TERMS X-ray capillary optics, digital mammography, scatter fraction, contrast improvement, MTF, breast cancer		15. NUMBER OF PAGES 15		
		16. PRICE CODE		
17. SECURITY CLASSIFICATION OF REPORT Unclassified	18. SECURITY CLASSIFICATION OF THIS PAGE Unclassified	19. SECURITY CLASSIFICATION OF ABSTRACT Unclassified	20. LIMITATION OF ABSTRACT Unlimited	

GENERAL INSTRUCTIONS FOR COMPLETING SF 298

The Report Documentation Page (RDP) is used in announcing and cataloging reports. It is important that this information be consistent with the rest of the report, particularly the cover and title page. Instructions for filling in each block of the form follow. It is important to *stay within the lines* to meet optical scanning requirements.

Block 1. Agency Use Only (Leave blank).

Block 2. Report Date. Full publication date including day, month, and year, if available (e.g. 1 Jan 88). Must cite at least the year.

Block 3. Type of Report and Dates Covered. State whether report is interim, final, etc. If applicable, enter inclusive report dates (e.g. 10 Jun 87 - 30 Jun 88).

Block 4. Title and Subtitle. A title is taken from the part of the report that provides the most meaningful and complete information. When a report is prepared in more than one volume, repeat the primary title, add volume number, and include subtitle for the specific volume. On classified documents enter the title classification in parentheses.

Block 5. Funding Numbers. To include contract and grant numbers; may include program element number(s), project number(s), task number(s), and work unit number(s). Use the following labels:

C - Contract	PR - Project
G - Grant	TA - Task
PE - Program Element	WU - Work Unit Accession No.

Block 6. Author(s). Name(s) of person(s) responsible for writing the report, performing the research, or credited with the content of the report. If editor or compiler, this should follow the name(s).

Block 7. Performing Organization Name(s) and Address(es). Self-explanatory.

Block 8. Performing Organization Report Number. Enter the unique alphanumeric report number(s) assigned by the organization performing the report.

Block 9. Sponsoring/Monitoring Agency Name(s) and Address(es). Self-explanatory.

Block 10. Sponsoring/Monitoring Agency Report Number. (If known)

Block 11. Supplementary Notes. Enter information not included elsewhere such as: Prepared in cooperation with...; Trans. of...; To be published in.... When a report is revised, include a statement whether the new report supersedes or supplements the older report.

Block 12a. Distribution/Availability Statement. Denotes public availability or limitations. Cite any availability to the public. Enter additional limitations or special markings in all capitals (e.g. NOFORN, REL, ITAR).

DOD - See DoDD 5230.24, "Distribution Statements on Technical Documents."

DOE - See authorities.

NASA - See Handbook NHB 2200.2.

NTIS - Leave blank.

Block 12b. Distribution Code.

DOD - Leave blank.

DOE - Enter DOE distribution categories from the Standard Distribution for Unclassified Scientific and Technical Reports.

NASA - Leave blank.

NTIS - Leave blank.

Block 13. Abstract. Include a brief (*Maximum 200 words*) factual summary of the most significant information contained in the report.

Block 14. Subject Terms. Keywords or phrases identifying major subjects in the report.

Block 15. Number of Pages. Enter the total number of pages.

Block 16. Price Code. Enter appropriate price code (*NTIS only*).

Blocks 17. - 19. Security Classifications. Self-explanatory. Enter U.S. Security Classification in accordance with U.S. Security Regulations (i.e., UNCLASSIFIED). If form contains classified information, stamp classification on the top and bottom of the page.

Block 20. Limitation of Abstract. This block must be completed to assign a limitation to the abstract. Enter either UL (unlimited) or SAR (same as report). An entry in this block is necessary if the abstract is to be limited. If blank, the abstract is assumed to be unlimited.

FOREWORD

Opinions, interpretations, conclusions and recommendations are those of the author and are not necessarily endorsed by the US Army.

Ch Where copyrighted material is quoted, permission has been obtained to use such material.

Ch Where material from documents designated for limited distribution is quoted, permission has been obtained to use the material.

Ch Citations of commercial organizations and trade names in this report do not constitute an official Department of Army endorsement or approval of the products or services of these organizations.

___ In conducting research using animals, the investigator(s) adhered to the "Guide for the Care and Use of Laboratory Animals," prepared by the Committee on Care and Use of Laboratory Animals of the Institute of Laboratory Resources, National Research Council (NIH Publication No. 86-23, Revised 1985).

___ For the protection of human subjects, the investigator(s) adhered to policies of applicable Federal Law 45 CFR 46.

___ In conducting research utilizing recombinant DNA technology, the investigator(s) adhered to current guidelines promulgated by the National Institutes of Health.

___ In the conduct of research utilizing recombinant DNA, the investigator(s) adhered to the NIH Guidelines for Research Involving Recombinant DNA Molecules.

___ In the conduct of research involving hazardous organisms, the investigator(s) adhered to the CDC-NIH Guide for Biosafety in Microbiological and Biomedical Laboratories.

Carolyn A MacDonnell 10/7/95
PI - Signature Date

TABLE OF CONTENTS

Page	
<u> 1 </u>	Front Cover
<u> 2 </u>	SF298
<u> 3 </u>	Foreword
<u> 4 </u>	Table of Contents
<u> 5 </u>	Introduction
<u> 6 </u>	Body
<u> 12 </u>	Conclusions
<u> 14 </u>	References
<u> 15 </u>	Appendix

INTRODUCTION

The quality of conventional mammograms has been limited in 1) contrast by the large fraction of scattered photons and the low inherent contrast of the breast tissue, and 2) in resolution by geometric unsharpness determined by the finite x-ray focal spot size. Also, in digital mammographic applications detector resolution has been a significant limitation. Computed radiography (CR) photostimulable phosphor plates have lower spatial resolution than film and have been limited in their application in mammography where the detection and evaluation of microcalcifications requires very high spatial resolution. Attempts to increase the effective resolution of the CR plates using conventional magnification techniques would be limited to modest magnifications due to the increase in blurring associated with the focal spot.

X-ray capillary optics^{1,2,3} small bundles of hollow capillary tubes, make use of the nearly total external reflection of x rays in a manner analogous to conventional fiber optics. These optics can be useful in addressing the problems above. X-ray capillary optics, due to their small angle of acceptance (critical angle = 1.6 milliradians for 20 keV photons), would reject at the entrance of the optic scattered photons which are at an angle greater than the critical angle. Focused arrays of these optics would be used between the breast and detector system to reduce scatter and enhance spatial resolution through magnification techniques.

At the onset of this postdoctoral fellowship, extensive measurements on the performance of individual capillaries and smaller non focused optics (at mammography energies) had been performed by Abreu et al.⁴. The fellowship allowed me to join the Medical Physics group at the University of Wisconsin at Madison where the imaging characteristics of capillary optics had begun to be investigated.

The purpose of the present work was to investigate the feasibility of using diverging glass capillary optics for digital mammography by modifying the scanning

gantry to permit measurements of the partial optics, performing measurements on the partial optic prototype in both static and scanned configurations and designing the final optics.

BODY

The mammographic scanning gantry was fabricated to allow the measurement of a partial capillary optic prototype built by X-Ray Optical Systems, Inc. (The scanning gantry allows the optic to be scanned in the horizontal plane while the phantom and detector shift in unison as shown in figure 1. The details of the scanner can be found in reference 5 which is a paper that our group submitted to Medical Physics.) The prototype optic is 16.7 cm long, and it tapers from an output diameter of 7.5 mm to an input diameter of 4.15 mm resulting in a magnification of 1.81. Though the prototype was not an optimal optic, it still was very useful in the performing of relevant imaging experiments.




Figure 1. Scanning gantry geometry. The optic is scanned in the horizontal plane while the detector and phantom are moved in parallel vertical planes to allow imaging of the wide area object.

First, the primary transmission of the partial optics prototype was measured in a static configuration. The primary transmission was approximately 46%. Scatter fractions with and without the optic were measured using a 5 cm thick Lucite phantom, a field size of 16 cm and a lead beam blocker with a width of 3.1 mm as shown in figure 2. A rectilinear scan was done to obtain an image for the scatter fraction measurement with the optic. Measurements were also done with and without a 5:1 scatter reducing grid. The results for these measurements are summarized in Table 1.




Figure 2. (a) Scatter fraction measurement geometry for normal acquisition with and without the grid. (b) The optic is scanned along with the lead shield. The shield is needed to keep scatter from bypassing the optic and being detected. The source to detector distance (SDD) = 42 cm and the source to phantom distance (SPD) = 24 cm.


Method	T _p	T _s	Scatter Fraction
Normal	~ 1	~1	0.450
5:1 Grid	0.660	0.200	0.170
Capillary Optic	0.460	0.003	0.018

Table 1. Scatter fractions and primary and scatter transmission factors of three methods of image acquisition.

From the scatter fractions in Table 1, a comparison was made of the capillary optic method to the normal method (no anti-scatter device used) to calculate the optic transmission necessary for equal SNR. That point was $\approx 50\%$. A comparison to the actual transmission of 46% shows that the optic gives the same SNR/dose as a 5:1 anti-scatter grid. Even though this optic is at the break even point compared to a grid, new manufacturing processes and greater linearity are expected to give higher primary x-ray transmissions. Any higher optic transmission will permit additional gains in SNR or a possible decrease in dose with its use.

The contrast improvement was also measured. This is important because it is directly related to reduced scatter. For these measurements a 5cm thick lucite contrast phantom containing holes of decreasing size in one direction and decreasing depth in another was used. To measure the contrast, the phantom was imaged, with grid and without a grid, using CR plates on a GE Senographe mammography unit set at 27 keV and 35 mAs. The phantom was then scanned using the capillary optics at 27 keV and 10 mAs. These images are shown in figure 3 and a quantitative contrast comparison of the cases with and without the grid are shown in figure 4.⁵ The capillary optic showed better performance compared to the cases with and without a grid for all hole sizes. The 5:1 grid yielded, an average contrast improvement factor of 1.2 compared to no grid. For the capillary optic the factor was 1.7, showing that the optic does an excellent job of reducing scattered radiation.

(a)



(b)




Figure 3




Figure 3. Images of contrast-detail phantom (a) without an anti-scatter grid, (b) with a 5:1 grid and (c) with the capillary optic.




Figure 4: The average measured contrast of the contrast-detail phantom is plotted versus depth for the methods with the optic, with a 5:1 grid and without a grid.

To determine the limiting resolution for the phosphor plate system with and without the optics, modulation transfer function (MTF) measurements were performed. The MTF was measured using the edge response function of the phosphor plate imaging system (CR) with and without the optic. (For details of these experiments, see reference 5.) The MTF curves are shown in figure 5. The MTF limiting resolution point for this comparison was defined to be the spatial frequency at which the MTF curve drops to 5%. When optimal geometric magnification (and large focal spot = 0.3 mm) is used the limiting resolution increases to 5.4 lp/mm. Using the capillary optic in a stationary mode (with a large focal spot = 0.3 mm), the 5% MTF level is increased to 9 lp/mm. With the scanned optic mode the limiting resolution decreased to 8 lp/mm. The scanned MTF is reduced because the scanning produces the possibility of blurring due to a non-linear taper of the optic or mere vibration of the scanning system. The performance of the scanned optic could be improved by having an optic with a more linear taper or by eliminating the vibration of the scanning system with the use of a more elaborate system.




Figure 5. MTFs of four CR imaging methods.

Currently, images of a Radiation Measurements, Inc. (RMI) Mammography Accreditation Phantom, Model 156, are being obtained with the scanning gantry and optic. Imaging a mammographic phantom that is used clinically with the optic is an important next step because it will allow us to subjectively evaluate the performance of the optic in phosphor plate mammography compared to conventional screen/film mammography, magnified screen/film mammography without the optics and magnified screen/film mammography with the optics. The RMI phantom contains objects which simulate micro-calcifications, fibrous structures in ducts, and tumor-like masses⁶. (A schematic of the RMI phantom is included in the appendix.) This phantom is used by mammography quality assurance groups as an integral test in determining the ability of the mammographic unit to image structures similar to those found clinically. Images of the phantom have been obtained in the normal fashion on film and on the CR plates with no optic. The screen/film combination does better, as expected, since the CR plates (without the optic) have a resolution limit of 5 lp/mm as opposed to the ~ 15 lp/mm resolution limit of the film. Since the optic has a small output diameter and the RMI phantom contains objects in a 8 cm x 8 cm region, the mammography unit was modified to allow it to run continuously so that the optic had enough time to scan the phantom. We are just starting to scan the phantom with the optic so there are no images available at this time.

CONCLUSIONS

Capillary optics showed nearly total rejection of scatter. This reduction in scattered radiation causes an improvement in image contrast by an average of 70% over methods with no anti-scatter device and 40% over methods using the 5:1 grid. The optic transmission was 46%, which is very promising since good total transmission is required to give a reasonable total breast dose. With the use of the optic (magnification

= 1.81), the limiting resolution of the CR system was increased to 9 lp/mm from 5 lp/mm without the optic. The resolution increase is linearly dependent on the magnification factor of the optic. If an optic with a magnification of 3 was obtained, then the limiting resolution could be increased to 15 lp/mm. New optics with larger magnifications should facilitate the use of phosphor plates as digital detectors capable of imaging the entire breast. Digital mammography would have all of the advantages of digital processing, such as image processing, image transmission and storage, and computer aided diagnosis.

The test optic was not large enough to be clinically feasible. However, because the measurements done on the optic showed excellent agreement with theoretical predictions, improvements in the manufacturing technology of the optics should produce imaging improvements as well.

Specific plans for next year:

After the images of the RMI phantom are obtained with the optic, noise power spectra measurements for the optic and the PCR plates will be set-up in the lab. We will investigate noise properties of the magnified phosphor plate images compared with conventional mammography.

In the next couple of months, we expect to receive (from X-Ray Optical Systems) an optic with a larger output diameter (up to 2 cm) and a more uniform linear taper. Once this optic is in hand, transmission, scatter, contrast and MTF measurements will be repeated to determine the overall imaging capabilities of the new optic. Hopefully, we can then image excised specimens and the RMI phantom with the new optic using the same modalities.

The scanner is also being redesigned so that only the optic/source will be scanned and the detector and object will remain motionless.

REFERENCES

- ¹ M.A. Khumakhov and V.Y. Shovkun, "Design of x-ray optical antiscatter grids for mammography", preprint IAE-5418/14. M., 1991, I.V. Kurchatov Institute of Atomic Energy.
- ² V.A. Arkadev, M.A. Khumakhov et al, "Wide-band x-ray optics with a large angular aperture", Sov. Physics. Usp. 32(3). March 1989.
- ³ M.A. Khumakhov and W. Gibson, "The Khumakhov Lens: A new optics for X-rays and neutrons." PIXAM conference of X-Ray Analysis, Honolulu, Hawaii, August 1991.
- ⁴ C. C. Abreu, D.G. Kruger, C.A. MacDonald, C.A. Mistretta, W.W. Peppler and Q.F. Xiao, "Measurements of capillary x-ray optics with potential for use in mammographic imaging," accepted to Med. Phys. tentative publication date Nov. 1995.
- ⁵ D.G. Kruger, C.C. Abreu, W.W. Peppler, C.A. MacDonald, C.A. Mistretta, "Imaging Characteristics of X-Ray Capillary Optics in Mammography," submitted, Medical Physics.
- ⁶ User Manual - Mammographic Accreditation Phantom, Model 156, RMI Radiation Measurements, Inc., Middleton, Wisconsin.

APPENDIX

1) Schematic of RMI phantom

	Region Materials
1.	1.56 mm nylon fiber
2.	1.12 mm nylon fiber
3.	0.89 mm nylon fiber
4.	0.75 mm nylon fiber
5.	0.54 mm nylon fiber
6.	0.40 mm nylon fiber
7.	0.54 mm simulated micro-calcification
8.	0.40 mm simulated micro-calcification
9.	0.32 mm simulated micro-calcification
10.	0.24 mm simulated micro-calcification
11.	0.16 mm simulated micro-calcification
12.	2.00 mm tumor-like mass
13.	1.00 mm tumor-like mass
14.	0.75 mm tumor-like mass
15.	0.50 mm tumor-like mass
16.	0.25 mm tumor-like mass

